

Bar



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204112	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 02/28/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Brookfield Police Department, 116 W. Brooks, Brookfield, MO		TIME OF INSPECTION 12:04 pm
CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.		
<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)		DATE AND TIME (from printout) 02/28/14 1204
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR	
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS	
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD	
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION	
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER	
<input checked="" type="checkbox"/> INDICATOR LIGHTS		
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 13001 EXP. DATE 03/07/2015		
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN G 6815 EXP. DATE 04/03/2014		
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)		
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)		
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE <input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE <input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE		
TEST 1 .097	TEST 2 .098	TEST 3 .098
<input checked="" type="checkbox"/> PERFORM R.F.I. TEST (PRINTOUT ATTACHED)		
INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)		
REFUSALS 0	(0-.04) 0	(.05-.09) 0
(.10-.14) 0	(.15-.19) 0	OVER .19 0
LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).		
Meets Department of Health Standards		
INSPECTING OFFICER		
SIGNATURE <i>Lane P. Quinn</i>		PRINT FULL NAME Lane P. Quinn
TYPE II PERMIT NUMBER/EXPIRATION DATE 240025 / 01-28-16		TELEPHONE NUMBER (660) 385-2132
RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office 2875 James Blvd. Poplar Bluff, MO 63901		

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13001

EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
LANE P QUINN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/28/2014

NUMBER 240025

EXPIRES 1/28/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

80-0771 (6-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator QUINN, LANE
Permit No 240025
Date Issued 1/28/2014 Date Expires 1/28/2016

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE LABORATORY
BAC DATAMASTER SERVICE NUMBER 240000
OFFICE 100

TESTING OFFICER:
QUINN, L. P.
OFFICER ID: 100
PERMIT NUMBER: 240000
EXPIRATION DATE: 01/26/10
MISCELLANEOUS DATA:

100 SUPERVISION BUREAU

BLANK TEST	1000	12:09
INTERNAL STANDARD	VERIFIED	12:09
EXTERNAL STANDARD	1007	12:09
BLANK TEST	1000	12:10
INTERNAL STANDARD	1000	12:10
BLANK TEST	1000	12:11
EXTERNAL STANDARD	1000	12:11
BLANK TEST	1000	12:12

100
100
100

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE LABORATORY
BAC DATAMASTER SERVICE NUMBER 240000
OFFICE 100

TESTING OFFICER:
QUINN, L. P.
OFFICER ID: 100
PERMIT NUMBER: 240000
EXPIRATION DATE: 01/26/10
MISCELLANEOUS DATA:

BLANK TEST 1000 12:12
INTERNAL STANDARD VERIFIED 12:12
EXTERNAL STANDARD 1000 12:12

Operator Signature

T.P. Quinn

Operator Signature

T.P. Quinn

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE PENITENTIARY
BAC DATAMASTER SERIAL NUMBER 100112
02/20/10
12/08

INFORMATION - CHECK

COMPUTERS OK

PROGRAMS 104-02-00000 OK

PRINTERS
SAMPLE CHAMBERS OK

FLOW DETECTORS OK

REPAIR
HIGH SPEED OK

DEVIATIONS OK

FILTERS OK

REPAIRS - STANDARD OK

REPAIRS - CALIBRATION OK

REMARKS - YES

01/20/10 10:00 AM - 01/20/10 10:00 AM - 01/20/10 10:00 AM
01/20/10 10:00 AM - 01/20/10 10:00 AM - 01/20/10 10:00 AM
01/20/10 10:00 AM - 01/20/10 10:00 AM - 01/20/10 10:00 AM

Operator Signature

TPR L.P. Davis